



#### KEY NOTES

1. SLAB SHALL BE RECESSED AT THE AUDIO BOOTH LOCATION IN ORDER TO COMPLY WITH ADA.
2. MOUNT A5245, HOOK, AT 2372 [4'-6"] AFF. PROVIDE 2-A5245'S FOR NAVY PROJECTS ONLY.  
MOUNT 1-A5145 AT 1372 [4'-6"] AFF AND 1-A5145 AT 1524 [5'-0"] AFF.
3. PROVIDE M4200, OTOSCOPE/OPHTHALMOSCOPE, FOR AIR FORCE AND NAVY PROJECTS.
4. M0025, AUDIOMETER, IS FOR NAVY PROJECTS AND IS OPTIONAL.
5. F3200, CLOCK, IS FOR NAVY PROJECTS AND IS OPTIONAL.
6. PROVIDE CD PLAYER AND VISUAL REINFORCEMENT ACCESSORIES FOR M0030, AUDIOMETER.
7. COORDINATE LOCATION OF M1620, PATIENT CHART HOLDER, AND ROOM SIGNAGE.
8. VERIFY THIS DIMENSION WITH MANUFACTURER'S VENTING REQUIREMENTS FOR THE M0041, AUDIO BOOTH.
9. COORDINATED ELECTRICAL REQUIREMENTS FOR AUDIO BOOTH WITH ACTUAL EQUIPMENT PROVIDED.

## PHYSICAL EVALUATION, HEARING SUITE, 2 ROOM (PEHS4)

AUDIOLOGY/SPEECH PATHOLOGY/ENT

FLOOR/EQUIPMENT PLAN

34.75 NSM (374 NSF)

DATE: 07/12/2001

0 1000 2000



1:50 SCALE (MILLIMETERS)